



Dimethylglyoxime, Powder

BAKER ANALYZED[®] A.C.S. Reagent

Product No. 1938

Lot No. K23622

Release Date 06/10/2011

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Meets A.C.S. Specifications		
Melting Point	238 - 241 °C.	240 °C.
Insoluble in Alcohol	0.05 % max.	0.01 %
Residue after Ignition	0.05 % max.	0.02 %
Suitability for Nickel Determination	Passes Test	Passes Test

For Laboratory, Research or Manufacturing Use

Country of Origin: INDIA

ISO

Phillipsburg, NJ 9001.2008, 14001.2004
 Paris, KY 9001.2008
 Mexico City, Mexico 9001.2008
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
 Selangor, Malaysia 9001.2008
 Panaji, India 9001.2008
 Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 1-800-582-2537 or 908-859-2151

Avantor[™] Performance Materials, Inc. (formerly Mallinckrodt Baker, Inc.)

222 Red School Lane • Phillipsburg, NJ 08865 • Phone: 908.859.2151 • Fax: 908.859.6905