



Diphenylamine, Crystal

BAKER ANALYZED[®] A.C.S. Reagent

Product No. 1944

Lot No. K28585

Release Date 07/11/2011

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Meets A.C.S. Specifications		
Meets Reagent Specifications for testing USP/NF monographs		
Melting Point	52.5 - 54.0 °C.	53.0 °C.
Solubility in Alcohol	Passes Test	Passes Test
Residue after Ignition	0.03 % max.	0.02 %
Nitrate (NO ₃)	Passes Test	Passes Test
Sensitivity for Nitrate	Passes Test	Passes Test

For Laboratory, Research or Manufacturing Use

Country of Origin: USA

ISO

Phillipsburg, NJ 9001.2008, 14001.2004
 Paris, KY 9001.2008
 Mexico City, Mexico 9001.2008
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
 Selangor, Malaysia 9001.2008
 Panaji, India 9001.2008
 Gliwice, Poland 9001.2008, 17025.2005


 Richard M. Siberski
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 1-800-582-2537 or 908-859-2151
 Avantor™ Performance Materials, Inc. (formerly Mallinckrodt Baker, Inc.)
 222 Red School Lane • Phillipsburg, NJ 08865 • Phone: 908.859.2151 • Fax: 908.859.6905