



# Sulfuric Acid, 1N Volumetric Solution

BAKER ANALYZED<sup>®</sup> Reagent

Product No. 5642

Lot No. L09522

Release Date 03/19/2012

## Certificate of Analysis

TEST	SPECIFICATION	RESULT
Standardization at 25°C traceable to NIST Standard Reference Material.		
SRM No		723e
Appearance	Passes Test	Passes Test
Normality	0.995 - 1.005	1.000
<b>Trace Impurities (in ppm):</b>		
Chloride (Cl)	1 max.	< 0.5
Phosphate (PO <sub>4</sub> )	1 max.	< 0.5
Heavy Metals (as Pb)	0.3 max.	< 0.3
Iron (Fe)	0.5 max.	< 0.3
For Laboratory, Research or Manufacturing Use		
<b>USE BEFORE March 31, 2014</b>		

Country of Origin: USA



Phillipsburg, NJ 9001.2008, 14001.2004  
 Paris, KY 9001.2008  
 Mexico City, Mexico 9001.2008  
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003  
 Selangor, Malaysia 9001.2008  
 Panaji, India 9001.2008  
 Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski  
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600  
 Avantor™ Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610