



Buffer Solution (Borate), pH 10 (Color Coded Blue)

BAKER ANALYZED[®] Reagent

Product No. 5655

Lot No. L07507

Release Date 02/27/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Each lot compared to NIST Standard Reference Buffer.		
pH at 25°C	9.99 - 10.01	10.00
Appearance	Passes Test	Passes Test
Product Information (not specifications):		
Potassium (K), Molarity (typical range)	0.094 - 0.114	
For Laboratory, Research or Manufacturing Use		
USE BEFORE February 28, 2014		
Country of Origin: USA		



Phillipsburg, NJ 9001.2008, 14001.2004
Paris, KY 9001.2008
Mexico City, Mexico 9001.2008
Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
Selangor, Malaysia 9001.2008
Panaji, India 9001.2008
Gliwice, Poland 9001.2008, 17025.2005


Richard M. Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor™ Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610