



Buffer Solution (Biphthalate), pH 4 (Color Coded Red)

BAKER ANALYZED[®] Reagent

Product No. 5657

Lot No. L10522

Release Date 03/13/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Each lot compared to NIST Standard Reference Buffer.		
pH at 25°C	3.96 - 4.04	4.01
Appearance	Passes Test	Passes Test
Product Information (not specifications):		
Potassium Biphthalate, Molarity		
(typical range)	0.0498 - 0.0502	
For Laboratory, Research or Manufacturing Use		
USE BEFORE March 31, 2014		

Country of Origin: USA



Phillipsburg, NJ 9001.2008, 14001.2004
Paris, KY 9001.2008
Mexico City, Mexico 9001.2008
Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
Selangor, Malaysia 9001.2008
Panaji, India 9001.2008
Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor™ Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610